



**URGENT PRODUCT DEFECT CORRECTION**

**AEROWIPE CLEANSING WIPES  
SUPPLIED IN VARIOUS FIRST AID KITS**

**ARTG 277597**

#	Batch Number	Expiry
<b>1</b>	24473	02/2025
<b>2</b>	24706	04/2025

Due to the visual presence of black spots resembling mould growth and following consultation with the Therapeutic Goods Administration, Aero Healthcare AU Pty Ltd is electing to recall two specific batches of AEROWIPE Cleansing Wipes supplied in some first aid kits purchased between **07/05/2020** and **25/08/2020**.

No other batches of AEROWIPE Cleansing Wipes are affected by this recall. A full lab investigation of this anomaly is in process.

Information about specific first aid kit models affected can be found here: [www.aerohealthcare.com/recalls](http://www.aerohealthcare.com/recalls)

If you have purchased a first aid kit supplied by Aero Healthcare between the above dates and it contains AEROWIPE Cleansing Wipes from batches 24473 or 24706, please **do not open** the wipes and call our customer service line to arrange for replacement wipes to be sent to you immediately.

**CUSTOMER SERVICE 1800 628 881**

Aero Healthcare sincerely regrets any inconvenience caused to their customers.